

Kanülierungsstandard THG Universitätsklinikum Frankfurt / Kardiotechnik Stand 2015

OP	Arteriell	Venös	CP	Vent	Retrograd (Option)	Hirnperfusion	Subclavia re	Y Adapter	Sternotomie	Körpertemperatur min
ACB	20 Fr	32/37	DLP 21012	15 Fr	RC2014S		Option		komplett	32
AKE	20 Fr	28/32	DLP 21012+Ostien	15 Fr			Option		POS	32
AKR	20 Fr	28/32	DLP 21012+Ostien	15 Fr			Option		POS	32
Tirone David	20 Fr	28/32	DLP 21012+Ostien	15 Fr			Option		POS	32
Ross Procedure	20 Fr	bicaval 22/22	DLP 21012+Ostien	15 Fr	RC2014S		Option		komplett	32
MKR	20 Fr	bicaval 22/22	DLP 21012	15 Fr	RC2014S		Option		POS	32
MKE	20 Fr	bicaval 22/22	DLP 21013	15 Fr	RC2014S		Option		POS	32
TKR	20 Fr	bicaval 22/22	DLP 21014	15 Fr	RC2014S		Option		POS	32
HTX	20 Fr	bicaval 22/22	DLP 21015				Option		komplett	32-34
ASD	20 Fr	bicaval 22/22	DLP 21016	15 Fr			Option		POS	34
VSD	20 Fr	bicaval 22/22	DLP 21017	15 Fr			Option		POS	34
Partieller Bogenersatz POS		28/32	DLP 21012+Ostien	15 Fr 2 x		14 Fr retrograd MIBB	18 / 20 Fr	X	POS	28
Bogenersatz kompl. Sternotomie		32/37	DLP 21012+Ostien	15 Fr 2 x	RC2014S	14 Fr retrograd MIBB	18 / 20 Fr	X	komplett	28
MKE/R+ACB	20 Fr	bicaval 22/22	DLP 21012	15 Fr	RC2014S		Option		komplett	32
AKE/R+ACB	20 Fr	32/37	DLP 21012+Ostien	15 Fr	RC2014S		Option		komplett	32
AKE/R+MKE/R	20 Fr	28/32	DLP 21012+Ostien	15 Fr	RC2014S		Option		POS	32
	ab 95 kg 24 Fr	ab 95 kg 22/26								
		< 65 kg 28/32								